

Tri Star Medical
 117 Lyle Lane
 Nashville, TN 37210
 paul@craftcannabistn.com
 650-444-7891

Sample: 07-18-2023-35901
 Sample Received: 07/18/2023;
 Report Created: 07/19/2023; Expires: 07/18/2024

Gas OG THCA Flower
 Plant, Flower - Cured



24.009 %

Total THC

0.293 %

Δ-9 THC

28.904 %

Total Cannabinoids

<LOQ %

Total CBD

Cannabinoids

Complete

(Testing Method: HPLC, CON-P-3000)
 Date Tested: 07/18/2023

Analyte	LOD	LOQ	Mass	Mass	
	%	%	%	mg/g	
Δ-8-Tetrahydrocannabinol (Δ-8 THC)	0.0493	0.0739	ND	ND	
Δ-9-Tetrahydrocannabinol (Δ-9 THC)	0.0493	0.0739	0.293	2.934	
Δ-9-Tetrahydrocannabinolic Acid (THCA-A)	0.0493	0.0739	27.041	270.414	
Δ-9-Tetrahydrocannabinophorol (Δ-9-THCP)	0.0493	0.0739	ND	ND	
Δ-9-Tetrahydrocannabivarin (Δ-9-THCV)	0.0493	0.0739	ND	ND	
Δ-9-Tetrahydrocannabivarinic Acid (Δ-9-THCVA)	0.0493	0.0739	0.145	1.445	
R-Δ-10-Tetrahydrocannabinol (R-Δ-10-THC)	0.0493	0.0739	ND	ND	
S-Δ-10-Tetrahydrocannabinol (S-Δ-10-THC)	0.0493	0.0739	ND	ND	
9R-Hexahydrocannabinol (9R-HHC)	0.0493	0.0739	ND	ND	
9S-Hexahydrocannabinol (9S-HHC)	0.0493	0.0739	ND	ND	
Tetrahydrocannabinol Acetate (THCO)	0.0493	0.0739	ND	ND	
Cannabidivarin (CBDV)	0.0493	0.0739	ND	ND	
Cannabidivarinic Acid (CBDVA)	0.0493	0.0739	ND	ND	
Cannabidiol (CBD)	0.0493	0.0739	ND	ND	
Cannabidiolic Acid (CBDA)	0.0453	0.0739	<LOQ	<LOQ	
Cannabigerol (CBG)	0.0493	0.0739	0.159	1.586	
Cannabigerolic Acid (CBGA)	0.0493	0.0739	1.142	11.419	
Cannabinol (CBN)	0.0493	0.0739	ND	ND	
Cannabinolic Acid (CBNA)	0.0493	0.0739	ND	ND	
Cannabichromene (CBC)	0.0493	0.0739	ND	ND	
Cannabichromenic Acid (CBCA)	0.0493	0.0739	0.124	1.241	
Total			28.904	289.039	

Total THC = THCa * 0.877 + Δ9-THC; Total CBD = CBDa * 0.877 + CBD; LOQ = Limit of Quantitation; ND = Not Detected.

Total THC Measurement of Uncertainty: ± 0.050%
 Total CBD Measurement of Uncertainty: ± 2.000%
 THCO potency analysis does not designate quantitative specificity of Δ-8-THCO and Δ-9-THCO isomers



New Bloom Labs
 6121 Heritage Park Drive, A500
 Chattanooga, TN 37416
 (844) 837-8223
 TN DEA#: RN0563975
 ANAB Testing Laboratory (AT-2868): ISO/IEC
 17025:2017

Natalie Siracusa
 Natalie Siracusa
 Laboratory Director

Powered by
 reLIMS
 info@relims.com